UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.usplo.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/523,993	05/23/2005	Yong-Nyun Kim	Q86186	2852
23373 SUGHRUE MI	7590 07/02/2007 ON PLIC		EXAM	INER
2100 PENNSY	LVANIA AVENUE, N.	w.	BOUCHELLE, LAURA A	
SUITE 800 WASHINGTO	N, DC 20037		ART UNIT	PAPER NUMBER
		•	3763	
•			MAIL DATE	DELIVERY MODE
			07/02/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	•	Application No.	Applicant(s)			
•		10/523,993	KIM, YONG-NYUN			
	Office Action Summary	Examiner	Art Unit			
	·	Laura A. Bouchelle	3763			
	The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply					
WHIC - Exter after - If NO - Failu Any	ORTENED STATUTORY PERIOD FOR REPLY CHEVER IS LONGER, FROM THE MAILING DANSIONS of time may be available under the provisions of 37 CFR 1.13 SIX (6) MONTHS from the mailing date of this communication. period for reply is specified above, the maximum statutory period were to reply within the set or extended period for reply will, by statute, reply received by the Office later than three months after the mailing ed patent term adjustment. See 37 CFR 1.704(b):	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be tin will apply and will expire SIX (6) MONTHS from a cause the application to become ABANDONE	N. nely filed the mailing date of this communication. D (35 U.S.C. § 133).			
Status						
2a)□	Responsive to communication(s) filed on <u>21 M</u> . This action is FINAL . 2b) This Since this application is in condition for allowar closed in accordance with the practice under E	action is non-final. nce except for formal matters, pro				
Disposit	ion of Claims					
5)□ 6)⊠ 7)□	Claim(s) 1-10 is/are pending in the application. 4a) Of the above claim(s) is/are withdraw Claim(s) is/are allowed. Claim(s) 1-10 is/are rejected. Claim(s) is/are objected to. Claim(s) are subject to restriction and/or	wn from consideration.				
Applicat	ion Papers					
9) The specification is objected to by the Examiner. 10) The drawing(s) filed on is/are: a) accepted or b) objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
Priority (under 35 U.S.C. § 119					
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 						
2) Noti	nt(s) ce of References Cited (PTO-892) ce of Draftsperson's Patent Drawing Review (PTO-948) rmation Disclosure Statement(s) (PTO/SB/08) er No(s)/Mail Date	4) Interview Summary Paper No(s)/Mail D 5) Notice of Informal F 6) Other:	ate			

Application/Control Number: 10/523,993 Page 2

Art Unit: 3763

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 1, 3, 4, 6-8 are rejected under 35 U.S.C. 102(b) as being anticipated by French et al (US 5131387). French discloses a moisture trap attached to a tube comprising a passage 66, an inlet port 34, a liquid absorption member 86 surrounding the passage, a gas permeable and liquid impermeable filter 102 for blocking the passage at a position father than the liquid absorption member from the inlet port (Col. 5, lines 60-65, Col. 7, lines 43-47). See Fig. 3. The device comprises an exhaust hole 48. The device comprises an extension 42 that isolates the absorption member and the filter from each other. See Fig. 2. The device provides a flow control (Col. 6, lines 2-4).

Claim Rejections - 35 USC § 103

- 3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
 - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Application/Control Number: 10/523,993 Page 3

Art Unit: 3763

4. Claims 5, 9, 10 are rejected under 35 U.S.C. 103(a) as being unpatentable over French in

view of Bell (US 5125415).

5. Claim 5 differs from French in calling for the filter to be made of a porous plastic resin

material. Bell teaches a cap having a filter made of a porous plastic resin material (Col 2, lines

57-58). It would have been obvious to one of ordinary skill in the art at the time of invention to

modify the device of French to have a filter formed of a porous plastic resin material as taught by

Bell because it is well known in the art to use a porous plastic resin filter to allow the passage of

gas and block the passage of fluid.

6. Claim 9 differs from French in calling for the cap to be attached to a liquid medicine

reservoir having a pressure device. Bell teaches a syringe having a liquid medicine reservoir and

a pressure device that allows the air to be purged from a syringe while preventing fluid from

escaping. Therefore, it would have been obvious to one of ordinary skill in the art at the time of

invention to modify the device of French to include a syringe as taught by Bell.

Response to Arguments

7. Applicant's arguments with respect to claims 1-10 have been considered but are moot in

view of the new ground(s) of rejection.

Application/Control Number: 10/523,993

Art Unit: 3763

Conclusion

Any inquiry concerning this communication or earlier communications from the

examiner should be directed to Laura A. Bouchelle whose telephone number is 571-272-2125.

The examiner can normally be reached on Monday-Friday 8-4.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's

supervisor, Nicholas Lucchesi can be reached on 517-272-4977. The fax phone number for the

organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent

Application Information Retrieval (PAIR) system. Status information for published applications

may be obtained from either Private PAIR or Public PAIR. Status information for unpublished

applications is available through Private PAIR only. For more information about the PAIR

system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR

system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would

like assistance from a USPTO Customer Service Representative or access to the automated

information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

TECHNOLOGY

Laura A Bouchelle

Page 4

Art Unit 3763